

AMENDMENT No. 3
to Agreement Ref. No. DC-3.5.1/13 dated 08.04.2020

Kyiv

28.12., 2020

THIS AMENDMENT to Agreement Reference No. DC-3.5.1/13 dated 08.04.2020 (together with all Annexes and Amendments hereto, this “Agreement”) is entered into between the **MINISTRY OF HEALTH OF UKRAINE** (the “MOH”), within the framework of the Loan Agreement for Serving People, Improving Health Project between the Government of Ukraine (the “Government”) and the International Bank for Reconstruction and Development, and the **UNITED NATIONS DEVELOPMENT PROGRAMME**, a subsidiary organ of the United Nations, an intergovernmental organization established by its Member States under the Charter signed on 26 June 1945, with its headquarters at 1 UN Plaza in New York, New York, 10017, USA (“UNDP” or the “UN Partner”, together with the Government, the “Parties” and each a “Party”).

NOW the Parties agree as follows:

1. To amend Clause 4 of the *Form of the Agreement* as follows:

“All activities under this Agreement shall be fully completed and all expenses incurred by June 30, 2021 (the “Completion Date”). The Completion Date can’t exceed the Project Closing date. UNDP shall issue the final financial statement not later than three (3) months after the Completion Date.”

2. To amend *Table III. Work Plan and Timeline of Annex I: Outputs and workplan* of the Agreement (as amended by Amendment No. 2 dated 30.09.2020 to the Agreement) as follows:

| N° | Activity | Months | | | | |
|----|---|--------------|---------------|------------|------------|---|
| | | January 2021 | February 2021 | March 2021 | April 2021 | Full Completion |
| 1 | Procurement and delivery of medical equipment and consumables | X | X | X | X | |
| | Progress Reports (per Annex III frequency) | X | X | X | X | |
| | Final Financial Statement | | | | | Within 3 months after Agreement Completion Date |

3. To amend *Annex VI: Specifications and Delivery Terms of the Items to Be Procured* of the Agreement, as specified in Annex to this Amendment.
4. All other Terms and Conditions of the Agreement remain unchanged.
5. This Amendment becomes effective on the date of its last signature.

IN WITNESS WHEREOF, the Parties hereto have executed this Amendment.

MINISTRY OF HEALTH OF UKRAINE



Name: Svitlana SHATALOVA
Title: Deputy Minister of Health of Ukraine

UNITED NATIONS DEVELOPMENT PROGRAMME



By:
Name: Dasha GERCHENKO
Title: Resident Representative, UNDP Ukraine

Annexto Amendment No. 3 dd. 28.12.2020

to Agreement Ref. No. DC-3.5.1/13

dd. 08.04.2020

ANNEX VI**SPECIFICATIONS AND DELIVERY TERMS OF THE ITEMS TO BE PROCURED**

| Activities | Description/Specification | Quantity, units | Price per unit, USD | Delivery terms |
|---|--|--------------------|---------------------------------------|--|
| 1. Procurement and delivery of medical equipment and consumables | | | | |
| 1.1. Refrigerator, laboratory | <p>Product description: upright refrigerator for storage of chemicals and reagents in clinical laboratory</p> <p>Certification:</p> <ul style="list-style-type: none"> ▪ Meeting quality standard for Manufacturer: ISO 13485 and ISO 9001 ▪ Declaration of Conformity as per current Ukrainian legislation <p>Product technical requirements:</p> <ol style="list-style-type: none"> 1. Refrigerator internal volume (useful capacity): 400L (+/-15%) 2. Temperature range °C 2°C / 8°C or wider and preset at +4°C/+5°C 3. Electronic temperature control. Accuracy, regardless of the load: +/- 1°C or better. 4. Compression type, CFC and HCFC Free Refrigerant. 5. Insulation material: polyurethane. 6. Fan-cooled or direct cooling for the distribution of air in the cabinet. 7. Interior chamber material - stainless steel structure/equivalent or Al plate. 8. Exterior cabinet material - painted steel / stainless steel. Door - solid (stainless steel or glass) | 60 | Estimated price per unit: 1,595.00 | Estimated term of first delivery: by February 28, 2021 |

| | | | | |
|--|---|--|--|--|
| | <p>9. Ambient operating temperature acceptable range: Lower 16°C, upper 32 °C.</p> <p>10. Power requirements: 220-230 V / 50-60 Hz.</p> <p>11. Power consumption: approx. 365 to 600 W (energy efficiency will be advantage).</p> <p>12. Supplied with Automatic voltage regulator</p> <p>13. Temperature monitoring (Display/Alarm functions):</p> <ul style="list-style-type: none"> ▪ External digital display with actual interior temperature, minimal graduation 1°C or better. ▪ Electronic temperature recording device includes data logger. ▪ Audio and visual alarm system indicates unsafe temperatures. ▪ Open door alarm. ▪ Battery back-up for alarm systems, and temperature recording device. <p>14. Accessories:</p> <ul style="list-style-type: none"> ▪ Door lock (key or electronic). ▪ Fitted with integrated castors. ▪ Shelves - easily adjustable shelves ▪ Any other accessories which is required for installation and operation should be included and described on the offered specification. <p>15. Transfer of ownership after delivery: after respective acts are signed by the MoH.</p> <p>16. Warranty: refrigerator is to be covered by a two-year replacement/maintenance warranty in the event of any component failure arising from defective design, materials or workmanship. Warranty obligations shall start after the transfer of ownership (after commissioning). <i>During warranty period the supplier shall provide free of charge two</i></p> | | | |
|--|---|--|--|--|

| | | | | |
|--|---|--|--|--|
| | <p><i>annual on-site technical support and maintenance visits.</i></p> <p>17. Materials: Restricted materials - the product and its constituent components, does not contain lead (except in batteries), mercury, cadmium, hexavalent chromium, polybrominated biphenyls (PBB) or polybrominated biphenyl ethers (PBDE). Other restricted materials: The appliance and its constituent components, does not contain lead, mercury, cadmium, hexavalent chromium, polybrominated biphenyls (PBB) or polybrominated biphenyl ethers (PBDE).</p> <p>18. Instructions for use: Refrigerator can be used in wide range of laboratory services from basic laboratory to reference center. Instructions for use, preventive maintenance and troubleshooting are supplied in English as a standard with the Russian/Ukrainian instructions available.</p> <p>19. Packaging and Labelling: Refrigerator will be packed with suitable protective packaging for the international shipment such as; packing foam, corrugated carton, non-fumigation pallets, pallet straps. The packaging will be compliant to restricted material requirements mentioned above.</p> <p>20. Installation and personnel training: All equipment shall be installed by an engineer with certification obtained from the manufacturer of appropriate equipment in accordance with technical requirements pertaining to equipment installation.</p> | | | |
|--|---|--|--|--|

| | | | | |
|---|--|----|--|--|
| | <p>The manufacturer must be presented officially in Ukraine by a distributor or a local agent. Customer personnel training in equipment operation will be provided by an engineer with certification obtained from the manufacturer of appropriate equipment.</p> | | | |
| 1.2. Ultra-low temperature laboratory freezer | <p>Product description: upright freezer for storage of chemicals and reagents in clinical laboratory</p> <p>Certification:</p> <ul style="list-style-type: none"> ▪ Meeting quality standard for Manufacturer: ISO 13485 and ISO 9001 ▪ Declaration of Conformity for medical devices as per Technical regulations (Decree of CMU #753 dd. 02.10.2013) <p>Product technical requirements:</p> <ol style="list-style-type: none"> 1. Temperature range: -70°C-80°C 2. Temperature control system 3. Display 4. Upright model design 5. Refrigerator internal volume (useful capacity): minimum 300 L 6. Alarm systems (Audio/visual): <ul style="list-style-type: none"> ▪ high / low temperature alarm ▪ power cut alarm ▪ open door alarm ▪ alarm of controller disfunction 7. Key lock 8. Equipped with shelves and boxes for long-term storage of test tubes with biological material. 9. Transfer of ownership after delivery: after respective acts are signed by the MoH. 10. Warranty: freezer is to be covered by a two-year replacement/maintenance warranty in the event of any component failure arising from defective design, materials or | 60 | Estimated price per unit: 10,000.00 | Estimated term of first delivery: by March 30, 2021 |

| | | | | |
|--|---|--|--|--|
| | <p>workmanship. Warranty obligations shall start after the transfer of ownership (after commissioning). <i>During warranty period the supplier shall provide free of charge two annual on-site technical support and maintenance visits.</i></p> <p>11. Materials: Restricted materials - the product and its constituent components, does not contain lead (except in batteries), mercury, cadmium, hexavalent chromium, polybrominated biphenyls (PBB) or polybrominated biphenyl ethers (PBDE). Other restricted materials: The appliance and its constituent components, does not contain lead, mercury, cadmium, hexavalent chromium, polybrominated biphenyls (PBB) or polybrominated biphenyl ethers (PBDE).</p> <p>12. Instructions for use: Freezer can be used in wide range of laboratory services from basic laboratory to reference center. Instructions for use, preventive maintenance and troubleshooting are supplied in English as a standard with the Russian/Ukrainian instructions available.</p> <p>13. Packaging and Labelling: Refrigerator will be packed with suitable protective packaging for the international shipment such as; packing foam, corrugated carton, non-fumigation pallets, pallet straps. The packaging will be compliant to restricted material requirements mentioned above.</p> <p>14. Installation and personnel training: All equipment shall be installed by an engineer with certification obtained from the manufacturer of appropriate equipment in</p> | | | |
|--|---|--|--|--|

| | | | | |
|--|--|--|--|--|
| | <p>accordance with technical requirements pertaining to equipment installation.</p> <p>15. The manufacturer must be presented officially in Ukraine by a distributor or a local agent.</p> <p>16. Customer personnel training in equipment operation will be provided by an engineer with certification obtained from the manufacturer of appropriate equipment.</p> | | | |
|--|--|--|--|--|

MINISTRY OF HEALTH OF UKRAINE

By: 


Name: Svitlana SHALOVA

Title: Deputy Minister of Health of Ukraine

UNITED NATIONS DEVELOPMENT PROGRAMME

By: 


Name: Daria GERCHENVA

Title: Resident Representative, UNDP Ukraine